

FLOW CYTOMETRY SELECTION GUIDE

Fluorophore	Laser	Ex (nm)	Em (nm)	Brightness
mFluor™ UV375	350	351	387	6
mFluor™ UV460	350	358	456	4
mFluor™ Violet 450	405	406	445	6
mFluor™ Violet 500	405	410	501	5
mFluor™ Violet 510	405	412	505	5
mFluor™ Violet 540	405	402	535	2
mFluor™ Violet 610	405	421	613	4
Alexa Fluor® 488	488	499	520	7
Fluorescein	488	498	517	6
iFluor® 488	488	491	516	8
mFluor™ Blue 590	488	569	589	3
mFluor™ Blue 620	488	589	616	3
PE	488	566	574	10
PE-Cy5	488	565	666	9
PE-Cy5.5	488	565	671	9
PE-Cy7	488	566	778	9
PE-iFluor® 594	488	566	606	9
PE-iFluor® 750	488	566	778	9
PE-Texas Red®	488	567	615	9
mFluor™ Green 620	532	525	623	1
TRITC	532	540	565	7
Cy3	561	554	568	4
mFluor™ Yellow 630	561	570	632	1
Texas Red®	561	586	603	4
APC	640	651	660	9
APC-Alexa Fluor® 750	640	650	774	7
APC-Cy7	640	651	779	7
APC-iFluor® 750	640	651	780	8
Cy5	640	650	669	7
Cy5.5	640	683	703	5
Cy7	640	754	778	2
Indocyanine Green	640	788	813	1
mFluor™ Red 700	640	680	695	3
mFluor™ Red 780	640	629	767	2

* Brightness scale is 1 (dimtest) to 10 (brightest), as a comparison of all included dyes. Brightness is computed as a function of extinction coefficient and quantum yield. Actual performance will vary due to instrumentation settings. Guide above was developed on a Cytex flow cytometer.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. © 2023 AAT Bioquest, Inc. All rights reserved. All trademarks and registered trademarks mentioned herein are the property of their respective owners. All other trademarks and registered trademarks are the property of AAT Bioquest.



www.stratech.co.uk
+44 (0) 1638 782600
orders@stratech.co.uk

 @stratech_uk
 @stratechscientifictd
 @stratech-scientific-ltd